

INDIAN INSTITUTE OF TECHNOLOGY KANPUR
Department of Earth Sciences

*** Revised**

Enquiry No: ES/IS/2014-2015/21

Date: 10.12.2014

Subject: Quotation for supply of Dust Sampler as mentioned below.

With reference to the subject mentioned above, you are invited to submit the quotation in a sealed cover in order to reach us by December 18, 2014 in the form of a hard copy to the address mentioned below. If you have any questions please call Dr. I. S. Sen at 2596440, email: isen@iitk.ac.in.

Technical Specifications for Dust Sampler:

- The instrument is required to collect the ambient particulate matter of aerodynamic size less than 10 micron and coarser particles separately in a cyclone collector.
- The cyclonic separation technique should be designed in order to have sharper cut off as close to 10 microns.
- The respirable dust sampler flow should be measured across a critical orifice and directly in m³/minute. Its flow metering calibration setups should be traceable to National standards.
- The instrument should have low noise and negligible impact of voltage fluctuations.
- The sampler should feature an automatic flow controller, which switches off, if flow rate is reduced beyond the tolerance level due to filter loading.
- The instrument should be designed in a way to attach with a gaseous sampler for monitoring gaseous pollutants like SO₂, NO_x.
- The instrument should be designed in such a manner that insures user's safety and remains unaffected from environmental factors (such as dust and temperature, air pressure). The instrument should therefore be rugged and reliable.
- The instrument should be easy to maintain.
- The vendor is also required to provide the filter papers along with the respirable suspended particulate matter sampler :
 - ✓ The size of the filter paper should be near about 8"x10" inches.
 - ✓ The filter paper should have low background for interference-free chemical determinants.
 - ✓ The filter paper should have low tare mass for accurate gravimetric analysis.
 - ✓ The pore size of the filter paper should be near about 1.5 -2.0 μm.
 - ✓ The filter paper should be heat treated to remove organic traces.
 - ✓ Required 200 sealed packets of filter paper.

Terms and Conditions:

- 1) Maximum education discount, if any should be offered.
- 2) Validity of quotation should be at least for 60 days.
- 3) Prices should be quoted on FOB (Free on Board) only (if imported).
- 4) Normal payment terms for the Institute will be applicable (90% on delivery of the items and the remaining 10% after satisfactory installation/ inspection).
- 5) Quotation should carry proper certifications like agency certificate, proprietary certificate, etc.

Kindly write the inquiry number on the top of envelop.

Dr.I.S.sen
WLE 201,
Earth Sciences Department
I.I.T. Kanpur PIN 208016